

# LZI SPICE II (UR-144/XLR-11) Calibrators – EU Only [IVD] For In Vitro Diagnostic Use Only



## Lin-Zhi International, Inc.

REF	Description	Quantity
0002c	THC Negative Calibrator	1 x 5 mL
0512	SPICE II (UR-144-N-(5-hydroxypentyl) metabolite) 10 ng/mL Calibrator	1 x 5 mL
0513	SPICE II (UR-144-N-(5-hydroxypentyl) metabolite) 20 ng/mL Calibrator	1 x 5 mL
0514	SPICE II (UR-144-N-(5-hydroxypentyl) metabolite) 35 ng/mL Calibrator	1 x 5 mL
0515	SPICE II (UR-144-N-(5-hydroxypentyl) metabolite) 50 ng/mL Calibrator	1 x 5 mL

### Intended Use

The Lin-Zhi International, Inc. (LZI) SPICE II (UR-144/XLR-11) Calibrators are for use as calibrators in the qualitative and semi-quantitative calibration of the LZI SPICE II (UR-144/XLR-11) Enzyme Immunoassay (Ref# 0510/0511) on a number of automated clinical chemistry analyzers.

### Description of the Calibrators:

The LZI SPICE II (UR-144/XLR-11) Calibrators are human urine-based liquids, and ready to use. The LZI THC Negative Calibrator is a processed drug-free human urine matrix containing buffers, stabilizers, and less than 0.1 % of sodium azide. The calibrators are prepared by spiking known concentrations of UR-144-N-(5-hydroxypentyl) metabolite into the LZI THC Negative Calibrator. Calibrators are made from NIST traceable standards.

\*Actual concentrations of these calibrators are within  $\pm 10\%$  of the target value as determined by GC/MS or LC/MS. Values are provided only as guidelines, and laboratories should determine the ranges based on their own test system and tolerance (2).

### Precautions and Warning

- The LZI SPICE II (UR-144/XLR-11) Calibrators are for in vitro diagnostic use only. Harmful if swallowed.
- The calibrators contain sodium azide, which may react with lead or copper plumbing to form potentially explosive metal azide. When disposing such liquids, always flush with a large volume of water to prevent azide build-up.
- The calibrators are prepared from non-sterile human urine. They are not tested by licensed reagents for the presence of antibodies to human immunodeficiency viruses, hepatitis antigens, and/or anti-hepatitis antibodies. They should be handled as potentially infectious. Always use good laboratory practice to avoid any skin contact or ingestion.
- Do not use the calibrators beyond their expiration dates.

### Preparation and Storage

The calibrators are ready-to-use. No reconstitution is required. Label the cap before removal to identify it with the original bottle. The calibrators should be stored refrigerated at 2-8°C when not in use.

### Stability

When stored refrigerated at 2-8°C, the calibrators are stable either opened-recapped or unopened until the expiration date printed on the vial label. Store calibrators tightly capped when not in use. Calibrator solution dispensed in the sample cups and left on board a clinical analyzer should be discarded after use.

### Procedure and Results

For qualitative calibration, use the 20 ng/mL as the cutoff calibrator. For semi-quantitative calibration, use all five calibrators. Recalibration should be performed after reagent bottle change, a change in calibrators or reagent lot, and after instrument maintenance is performed. For interpretation of results, refer to the appropriate LZI SPICE II (UR-144/XLR-11) Enzyme Immunoassay (Ref# 0510/0511) package insert (3).

### Limitations

The LZI SPICE II (UR-144/XLR-11) Calibrators are for use with the LZI SPICE II (UR-144/XLR-11) Enzyme Immunoassay (Ref# 0510/0511) to detect UR-144/XLR-11-type synthetic cannabinoids in human urine only.

### Bibliography

1. Urine testing for Drug of Abuse. National Institute on Drug Abuse (NIDA) Research Monograph 73, 1986.
2. Guidance for Industry, Abbreviated 510(k) Submissions for In Vitro Diagnostic Calibrators. U.S. Department of Health and Human Services. FDA, Document issued on February 22, 1999.
3. LZI SPICE II (UR-144/XLR-11) Enzyme Immunoassay (Ref# 0510/0511) package insert.

**Notice: Adulteration of reagents, use of instruments without appropriate capabilities, or other failure to follow instructions as set forth in this labeling can affect performance characteristics, and stated or implied label claims.**



**Manufacturer:**  
Lin-Zhi International, Inc.  
2945 Oakmead Village Court  
Santa Clara, CA 95051  
USA  
Tel: (408) 970-8811  
Fax: (408) 970-9030  
www.lin-zhi.com



### Authorized European Rep. within the EU:

CEpartner4U  
Esdoornlaan 13  
3951 DB Maarn  
The Netherlands  
www.cepartner4u.eu

