

LZI Oral Fluid Methadone Enzyme Immunoassay

REF S0110c (75/37.5 mL R₁/R₂ Kit)
S0111c (750/375 mL R₁/R₂ Kit)



FOR RESEARCH & DEVELOPMENT USE ONLY

Lin-Zhi International, Inc.

Intended Use

The Lin-Zhi International, Inc. (LZI) Oral Fluid Methadone Enzyme Immunoassay is a homogeneous enzyme immunoassay intended for the qualitative and semi-quantitative determination of methadone in neat human oral fluid, collected into an LZI Oral Fluid Collector, at the cutoff value of 20 ng/mL. The assay is designed for prescription use with a number of automated clinical chemistry analyzers. This is a Non-FDA Approved assay for Research & Development use only and as such should not be repackaged for in vitro diagnostic use.

The assay provides a rapid screening procedure for determining the presence of methadone in human oral fluid. The assay provides only a preliminary analytical result. A more specific alternative analytical chemistry method must be used in order to obtain a confirmed analytical result. Gas or Liquid Chromatography/Mass Spectrometry (GC/MS or LC/MS) is the preferred confirmatory method (1, 2). Clinical consideration and professional judgment should be exercised with any drug of abuse test result, particularly when the preliminary test result is positive.

Summary and Explanation of Test

Methadone is a synthetic diphenylheptanonylamine opioid that has similar analgesic activity and potency to morphine when administered parentally. However, unlike morphine, it reliably retains its effectiveness when given orally, and tolerance and physical dependency develop slowly (3, 4). Although methadone is prescribed to relieve chronic pain, its primary medical application is the detoxification and maintenance treatment of a narcotic or heroin addiction (3-6). The abuse potential of methadone is comparable to that of morphine due to its resembling pharmacological activity (3, 5, 7). Methadone is available in tablets and as a solution for parental injection. It is readily absorbed from the gastrointestinal tract when ingested, and metabolized extensively in the liver. The major metabolite is 2-ethylidene-1,5-dimethyl-3,3-diphenylpyrrolidine, commonly known as EDDP (8). The metabolites are secreted in urine or bile along with unchanged drug. Both parent drug and its metabolite are diffused into oral fluid from the blood stream. Detection of methadone and its metabolite in the oral fluid is an indication of methadone use.

Due to the rapid metabolism of methadone, the assay is well suited for the measurement of methadone treatment compliance, and highlights the use of illicit methadone as well as the consumption of methadone in ways other than prescribed (9).

Assay Principle

The LZI Oral Fluid Methadone Enzyme Immunoassay is a homogeneous enzyme immunoassay with ready-to-use liquid reagents. The assay is based on competition between drug in the sample and drug labeled with the enzyme glucose-6-phosphate dehydrogenase (G6PDH) for a fixed amount of antibody in the reagent (10). Enzyme activity decreases upon binding to the antibody, and the drug concentration in the sample is measured in terms of enzyme activity. In the absence of drug in the sample, methadone-labeled G6PDH conjugate is bound to antibody, and the enzyme activity is inhibited. When drug is present in the sample, antibody binds to free drug and the unbound methadone-labeled G6PDH then exhibits its maximal enzyme activity. Active enzyme converts nicotinamide adenine dinucleotide (NAD) to NADH, resulting in an absorbance change that can be measured spectrophotometrically at 340 nm.

Reagents Provided

Antibody/Substrate Reagent (R₁): Contains mouse monoclonal anti-methadone antibody, glucose-6-phosphate (G6P), nicotinamide adenine dinucleotide (NAD), stabilizers, and sodium azide (0.09 %) as a preservative.

Enzyme-drug Conjugate Reagent (R₂): Contains methadone-labeled glucose-6-phosphate dehydrogenase (G6PDH) in buffer with sodium azide (0.09 %) as a preservative.

Calibrators and Controls*

*Calibrators and Controls are sold separately and contain negative synthetic oral fluid matrix with Proclin 300™ as a preservative.

ORAL FLUID METHADONE Calibrator/Control	REF #
Oral Fluid LIS Negative Calibrator	S0008
Cutoff Calibrator: Contains 20 ng/mL methadone	S0113c
High Calibrator: Contains 50 ng/mL methadone	S0115c
Low Calibrator / Level 1 Control: Contains 10 ng/mL methadone	S0117c
Intermediate Calibrator / Level 2 Control: Contains 30 ng/mL methadone	S0118c

Collectors **

**Collectors are sold separately.

ORAL FLUID Collector	REF #
LZI Oral Fluid Collector -50 mL Polypropylene Centrifuge Tube	S0000b

Precautions and Warning

- This test is a Non-FDA approved assay and is for Research & Development use only. This test should not be repackaged for in vitro diagnostic use.
- Harmful if swallowed.
- Reagent contains sodium azide as a preservative, which may form explosive compounds in metal drain lines. When disposing such reagents or wastes, always flush with a large volume of water to prevent azide buildup. See National Institute for Occupational Safety and Health Bulletin: Explosive Azide Hazards (8/16/76).
- Do not use the reagents beyond their expiration dates.

Reagent Preparation and Storage

The reagents are ready to use. No reagent preparation is required. All assay components should be refrigerated at 2-8°C when not in use.

Specimen Storage and Shipping

Note: If oral fluid samples cannot be analyzed immediately, they may be stored in amber glass vials and refrigerated (2-8°C) for up to 21 days or frozen (-20°C) for up to 21 days. Studies have been performed up to 21 days to show methadone is stable in oral fluid. No further study was conducted beyond 21 days.

Samples should always be shipped cold (2-8°C), packed in gel ice, and shipped for next day delivery (within 24 hours). Failure to store or ship samples under these conditions may result in a significant decrease in recovery of analyte. Please see additional details in the Specimen Collection and Handling section below.

Specimen Collection and Handling

Oral fluid samples should be collected into a device without an absorbing pad, such as the LZI Oral Fluid Collector (a 50 mL polypropylene centrifuge tube) (11).

Prior to testing, samples should be frozen overnight (at minimum) and then allowed to thaw at room temperature. Samples should then be spun for five minutes at 3000 rpm to remove particulates. Only the clear top layer should be assayed for EIA testing and/or confirmatory testing. Samples should be at room temperature (18-25°C) for testing.

Samples do not require dilution or any additional correction factors.

Fresh and properly stored oral fluid samples should be within the normal pH range of 6-8; however, any sample with pH ranging from 3-10 can be tested without any pretreatment of the samples.

Handle all oral fluid specimens as if they are potentially infectious.

Instrument

Clinical chemistry analyzers capable of maintaining a constant temperature, pipetting sample, mixing reagents, measuring enzyme rates at 340 nm, and timing the reaction accurately can be used to perform this homogeneous enzyme immunoassay. Performance characteristics presented in this package insert have been validated on the Hitachi 717. If other instruments are used, performance will need to be validated by the laboratory.

Assay Procedure

Analyzers with specifications indicated above are suitable for performing this homogeneous enzyme immunoassay. Refer to the specific parameters used for each analyzer before performing the assay. Typical assay parameters used for the Hitachi 717 analyzer include a 60 µL sample, 150 µL of antibody reagent (R₁), and 75 µL of enzyme conjugate reagent (R₂), in 37°C incubation temperature, 30-35 reading frames, and 340 nm primary wavelength.

For qualitative analysis, use the 20 ng/mL cutoff calibrator. For semi-quantitative analysis, use all five calibrators and controls. Recalibration should be performed after a reagent bottle change or a change in calibrators or reagent lot. Two levels of controls are available for monitoring the cutoff level: 10 ng/mL and 30 ng/mL.

Calibration and Quality Control

Good laboratory practices recommend the use of at least two levels of control specimens (one positive and one negative control near the cutoff) to ensure proper assay performance. Controls should be run with each new calibration and after specific maintenance or troubleshooting procedures as detailed in the instrument system manual. Each laboratory should establish its own quality control frequency. If any trends or sudden change in control value are observed, review all operating parameters or contact LZI technical support for further assistance. Laboratories should comply with all federal, state, and local laws, guidelines, and regulations.

Results

Note: A positive test result does not always mean a person took illegal drugs and a negative test result does not always mean a person did not take illegal drugs. There are number of factors that influence the reliability of drug tests.

Qualitative: The cutoff calibrator, which contains 20 ng/mL of methadone, is used as a reference for distinguishing positive from negative samples. A sample with a change in absorbance ($\Delta\text{mA}/\text{min}$) equal to or greater than that obtained with the cutoff calibrator is considered positive. A sample with a change in absorbance ($\Delta\text{mA}/\text{min}$) lower than that obtained with the cutoff calibrator is considered negative.

Semi-Quantitative: The semi-quantitative mode is for purposes of (1) enabling laboratories to determine an appropriate dilution of the specimen for verification by a confirmatory method such as GC/MS or LC/MS, or (2) permitting laboratories to establish quality control procedures.

When an approximation of concentration is required, a calibration curve can be established with the five assay calibrators and controls. The concentration of methadone in the sample may then be estimated from the calibration curve.

Limitations

1. A positive result from the assay indicates only the presence of methadone. The test is not intended for quantifying this single analyte in samples.
2. A positive result does not necessarily indicate drug abuse.
3. A negative result does not necessarily mean a person did not take methadone.
4. There is a possibility that other substances and/or factors not listed above may interfere with the test and cause incorrect results (e.g., technical or procedural error, fluid intake, endogeneous or exogeneous interferences).
5. Positive results should be confirmed by other affirmative, analytical *chemistry* methods (e.g., chromatography), preferably GC/MS or LC/MS.
6. The test is designed for use with human oral fluid only.

Bibliography:

1. Urine Testing for Drug of Abuse, National Institute on Drug Abuse (NIDA) Research Monograph 73 (1986).
2. Mandatory Guidelines for Federal Workplace Drug Testing Program, National Institute on Drug Abuse, Federal Register, **69**(71): 19644 (2004).
3. Goodman, L.S., and Gilman, A.. The Pharmacological Basis of Therapeutics, 4th edition, 380, The MacMillan Co., (1970).
4. Roper, J.D., Garside, D., and Goldberger, B.A., "Opiates", in Contemporary Practice in Clinical Toxicology, 2nd edition, Leslie M. Shaw, editor-in-chief, AACCC, (2000).
5. Lacy, C., Armstrong, L.L., Lipsy, R.J., and Lance, L.L. Drug Information Handbook. Hudson, OH: Lexi-Comp, (1993).
6. Katzung, B.G., Basic and Clinical Pharmacology, 6th edition, Appleton & Lange, Norwalk, CT. (1995).
7. Physician's Desk Reference, 54th edition, Medical Economics Company, Montvale, NJ; 2711-2713 (2000).
8. Baselt, R.C. and Carvey, R.H., Disposition of Toxic Drugs and Chemicals in Man, 4th edition, Chemical Toxicology Institute, Foster City, CA. (1995).
9. Wolf, K. and Hay, A., "Methadone in Saliva", *Clin.Chem.*, **37**(7) (1991).
10. Rubenstein, K.E., Schneider, R.S., and Ullman, E.F., Homogeneous Enzyme Immunoassay: A New Immunochemical Technique, *Biochem Biophys Res Commun*, **47**, 846 (1972).
11. LZI Oral Fluid Sample Preparation Sheet.

For technical assistance please call: (408) 970-8811

Manufacturer:

 **Lin-Zhi International, Inc.**
2945 Oakmead Village Court
Santa Clara, CA 95051
USA
Tel: (408) 970-8811
Fax: (408) 970-9030
www.lin-zhi.com

© March 2017 Rev. 6