LZI Methadone Metabolite (EDDP) (300) Calibrators

IVD For In Vitro Diagnostic Use Only

2°C 8°C

Lin-Zhi International, Inc.

Outside USA Only

REF	Description	Quantity
0001	LZI Universal Negative Calibrator	1 x 5 mL
0192b	LZI Methadone Metabolite (EDDP) (300) Low Calibrator (150 ng/mL)	1 x 5 mL
0193b	LZI Methadone Metabolite (EDDP) (300) Cutoff Calibrator (300 ng/mL)	1 x 5 mL
0194b	LZI Methadone Metabolite (EDDP) (300) Intermediate Calibrator (600 ng/mL)	1 x 5 mL
0195b	LZI Methadone Metabolite (EDDP) (300) High Calibrator (1000 ng/mL)	1 x 5 mL

Intended Use

The Lin-Zhi International, Inc. (LZI) Methadone Metabolite (EDDP) (300) Calibrators are for use as calibrators in the qualitative and semi-quantitative calibration of the LZI Methadone Metabolite (EDDP) Enzyme Immunoassay (Ref# 0190b/0191b) at the cutoff value of 300 ng/mL on automated clinical chemistry analyzers (1).

Description of the Calibrators

The LZI Methadone Metabolite (EDDP) (300) Calibrators are human urine-based liquids and ready-to-use. The LZI Negative Calibrator is a processed drug-free human urine matrix containing buffers, stabilizers, and less than 0.1% of sodium azide. The calibrators are prepared by spiking known concentrations of 2-ethylidene-1,5-dimethyl-3,3-diphenylpyrrolidine (EDDP) into the LZI Negative Calibrator.

*****Actual concentrations of these calibrators are confirmed by GC/MS or LC/MS. Values are provided only as guidelines and laboratories should determine the ranges based on their own test system and tolerance (2).

Precautions and Warning

- *The LZI Methadone Metabolite (EDDP) (300) Calibrators are for in vitro diagnostic use only. Harmful if swallowed.*
- *The calibrators contain sodium azide, which may react with lead or copper plumbing to form potentially explosive metal azide. When disposing of such liquids always flush with a large volume of water to prevent azide build-up (3).*
- *The calibrators are prepared from non-sterile human urine. They are not tested by licensed reagents for the presence of antibodies to human immunodeficiency viruses, hepatitis antigens, and/or anti-hepatitis antibodies. They should be handled as potentially infectious. Always use good laboratory practices to avoid any skin contact or ingestion.*
- *Do not use the calibrators beyond their expiration dates.*

Preparation and Storage

The calibrators are ready-to-use. No reconstitution is required. Label the cap before removal to identify it with the original bottle. The calibrators should be stored refrigerated at 2-8ºC when not in use. See the expiration date on individual bottle labels.

Stability

When stored refrigerated at 2-8ºC, the calibrators are stable either opened-recapped or unopened, until the expiration date printed on the vial label. Store calibrators tightly capped when not in use. Calibrator solution dispensed in the sample cups and left onboard a clinical analyzer should be discarded after use.

Procedure and Results

For qualitative calibration, use the 300 ng/mL as the cutoff calibrator. For semi-quantitative calibration, use all five calibrators. Recalibration should be performed after reagent bottle change, a change in calibrators or reagent lot, and after instrument maintenance is performed. For interpretation of results, refer to the appropriate LZI Methadone Metabolite (EDDP) Enzyme Immunoassay (Ref# 0190b/0191b) package insert (4).

Limitations

The LZI Methadone Metabolite (EDDP) (300) Calibrators are for use with the LZI Methadone Metabolite (EDDP) Enzyme Immunoassay (Ref# 0190b/0191b) to detect Methadone Metabolite (EDDP) in human urine only.

Bibliography

1. Urine testing for Drug of Abuse. National Institute on Drug Abuse (NIDA) Research Monograph 73, 1986.
2. Guidance for Industry and Food and Drug Administration Staff, The Abbreviated 510(k) Program. U.S. Department of Health and Human Services. FDA, Document issued on September 13, 2019.
3. Sodium Azide. National Institute for Occupational Safety (NIOSH) Pocket Guide to Chemical Hazards. Available online at: https://www.cdc.gov/niosh/npg/npgd0560.html.
4. LZI Methadone Metabolite (EDDP) Enzyme Immunoassay (Ref# 0190b/0191b) package insert.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

Symbols

LZI uses the symbols and signs listed on the symbol glossary on the website. Visit www.lin-zhi.com/symbol-glossary for detailed information.

Additions, deletions, or changes are indicated by a change bar in the margin.
For technical assistance please call: (408) 970-8811

Notice: Adulteration of reagents, use of instruments without appropriate capabilities, or other failure to follow instructions as set forth in this labeling can affect performance characteristics, and stated or implied label claims.

Manufacturer:
Lin-Zhi International, Inc.
2945 Oakmead Village Court
Santa Clara, CA 95051
USA
Tel: (408) 970-8811
Fax: (408) 970-9030
www.lin-zhi.com

Authorized European Rep. within the EU:
AR Experts B.V.
Boeingavenue 209,
1119 PD Schiphol-Rijk,
The Netherlands

Importer within the EU:
MedEnvoy Global B.V.
Prinses Margrietplantsoen 33 – Suite 123
2595 AM The Hague,
The Netherlands

UK Responsible Person within the UK:
UKCA Experts Ltd.
6th Floor City Gate East,
Tollhouse Hill
Nottingham
NG1 5FS,
The United Kingdom (UK)

Importer within the UK:
MedEnvoy UK Limited
85 Great Portland Street,
First Floor, London
W1W 7LT,
The United Kingdom (UK)

© February 2026 Rev. 0